



太一電子檢測有限公司 校正實驗室
Tai Yi Electronics & Surveillance Co., Ltd. Calibration Laboratory

校正報告

Report of Calibration



Calibration Laboratory

1625

校正日期

2015/07/02

Calibration Date

儀器名稱

溫濕度記錄器

Equipment

廠牌

DICKSON

Manufacturer

型號

XH100

Model No.

序號/識別號碼

13087291

Serial No./ ID No.

送校單位

工業技術研究院

Applicant

送校單位地址

新竹縣竹東鎮中興路四段195號

Applicant Address

- 上項儀器經本實驗室以誠信的態度執行校正作業，校正結果詳述於本報告內。
- The above instrument has been calibrated by Tai Yi Laboratory with an honest attitude. The details of the calibration results can be found in this report.
- 本報告內之數值是在本實驗室規定之環境下執行校正所得的結果。
- The results in this report have been obtained from a controlled laboratory environment.
- 本報告校正之結果僅對校正報告內提及之送校件有效。
- These results are only effective for the UUT ("Unit Under Test"), written in this report.
- 本校正報告未得到實驗室書面同意不得任意摘錄或複製使用，但全文複製除外。
- This report shall not be reproduced in any form, except in full, without the prior written approval of the calibration laboratory.



校正報告專用章
Report Seal

報告簽署人
Signed by

張俊育

22205 新北市深坑區北深路三段270巷12號3樓(信義財貿中心)
3F., No.12, Ln. 270, Sec. 3, Beishen Rd., Shenkeng Dist., New Taipei City 22205, Taiwan (R.O.C.)
TEL: +886-2-26627199 FAX: +886-2-26626377
E-mail: servicetai@ty-es.com.tw http://www.ty-es.com.tw

Report No.: T1507010801





► 校正結果 Calibration Results

• 溫濕度量測 Temperature and Relative Humidity Measurement

標準值	器示值	器差值	擴充不確定度
20 °C	20 °C	0 °C	1 °C
40 %RH	46 %RH	6 %RH	3 %RH
25 °C	25 °C	0 °C	1 °C
60 %RH	65 %RH	5 %RH	3 %RH
30 °C	30 °C	0 °C	1 °C
80 %RH	82 %RH	2 %RH	3 %RH

• 實驗室依據ISO/IEC 17025之規定，不對校正週期及允收標準做出判定。使用者依自訂之校正週期按時送校。

• According to ISO/IEC 17025, a laboratory should neither suggest calibration period nor make judgments. Please follow the calibration schedule as set by the user.

Report No.: T1507010801

